



California Medical Device Recall Information



Recall Name

Respironics California Recalls V60 Ventilators Due to the Potential for the Ventilator to Shut Down Without Alarm

| Recall Date | Product Description | Recalling Firm | Recall Reason |
|--------------|--|---|---|
| 06/03/13 | V60 Ventilators containing Power Management Board Assemblies | Respironics California, Inc. [a Div. of Philips Healthcare] Carlsbad, CA | <i>Potential software issue in the V60 Power Management Board Assembly. If a Power Management Board Assembly component fails, it may cause ventilator support to be lost with no audible alarm.</i> |
| Recall Class | Product Identification | Distribution | Affected Dates |
| I | V60 Ventilators V60 Power Management Board Assemblies <u>Affected Material Part Numbers and Serial Numbers</u> | CA , nationwide | Shipped prior to 04/01/13 |

FOR ADDITIONAL INFORMATION, PLEASE VISIT:

<http://www.fda.gov/MedicalDevices/Safety/ListofRecalls/ucm358028.htm>